

Catalog No.: 0225 (100 mL Kit)  
0226 (500 mL Kit)

### Intended Use

The DRI® Barbiturate Assay is intended for the qualitative and semiquantitative determination of barbiturates in human urine.

***This assay provides only a preliminary analytical test result. A more specific alternative chemical method must be used in order to obtain a confirmed analytical result. Gas chromatography/mass spectrometry (GC/MS) is the preferred confirmatory method.<sup>1,2</sup> Clinical consideration and professional judgment should be applied to any drug of abuse test result, particularly when preliminary positive results are used.***

### Summary and Explanation of the Test

Drug abusers may abuse various barbiturates, such as short-acting secobarbital and long-acting phenobarbital, through oral ingestion or by intravenous and/or intramuscular injection. Long-term abuse can lead to respiratory depression or, in severe cases, coma. When ingested, a barbiturate is rapidly metabolized and excreted into urine, allowing immunoassays to detect recent use.

The DRI Barbiturate Assay is a homogeneous enzyme immunoassay<sup>3</sup> using ready-to-use liquid reagents. The assay uses monoclonal antibodies that detect most barbiturates in urine. The assay is based on the competition of an enzyme glucose-6-phosphate dehydrogenase (G6PDH) labeled drug and the drug from the urine sample for a fixed amount of specific antibody binding sites. In the absence of drug from the sample, the G6PDH labeled drug is bound by the specific antibody and the enzyme activity is inhibited. This phenomenon creates a relationship between drug concentration in urine and the enzyme activity. The enzyme G6PDH activity is determined spectrophotometrically at 340 nm by measuring its ability to convert nicotinamide adenine dinucleotide (NAD) to NADH.

### Reagents

**Antibody/Substrate Reagent.** Contains monoclonal anti-barbiturate antibody, glucose-6-phosphate (G6P), and nicotinamide adenine dinucleotide (NAD) in Tris buffer with sodium azide as preservative.

**Enzyme Conjugate Reagent.** Contains barbiturate labeled with glucose-6-phosphate dehydrogenase (G6PDH) in Tris buffer with sodium azide as preservative.

### Additional Materials Required (sold separately):

Catalog No. 1664 DRI Negative Calibrator; 10 mL  
1388 DRI Negative Calibrator; 25 mL  
1588 DRI MultiDrug Calibrator 1, 10 mL  
1589 DRI MultiDrug Calibrator 1, 25 mL  
1591 DRI MultiDrug Calibrator 2, 10 mL  
1592 DRI MultiDrug Calibrator 2, 25 mL  
1594 DRI MultiDrug Calibrator 3, 10 mL  
1595 DRI MultiDrug Calibrator 3, 25 mL  
1597 DRI MultiDrug Calibrator 4, 10 mL  
1598 DRI MultiDrug Calibrator 4, 25 mL  
1599 DRI MultiDrug Urine Control 1, 10 mL  
1553 DRI MultiDrug Urine Control 1, 25 mL  
1600 DRI MultiDrug Urine Control 2, 10 mL  
1555 DRI MultiDrug Urine Control 2, 25 mL

### Precautions and Warnings

1. This test is for in vitro diagnostic use only. The reagents are harmful if swallowed.
2. Reagents used in the assay components contain sodium azide, which may react with lead or copper plumbing to form potentially explosive metal azides. When disposing of such reagents, always flush with a large volume of water to prevent azide build-up.
3. Do not use the reagents beyond their expiration dates.

### Reagent Preparation and Storage

The reagents are ready for use. No reagent preparation is required. All assay components, when stored at 2-8°C, are stable until the expiration date indicated on the label.

### Specimen Collection and Handling

Collect urine specimens in plastic or glass containers. Testing of fresh urine specimens is suggested.

*The Mandatory Guidelines for Federal Workplace Drug Testing Programs; Final Guidelines* recommends that specimens that do not receive an initial test within 7 days of arrival in the laboratory should be placed into secure refrigeration units.

Samples within a pH range of 3 to 11 are suitable for testing with this assay.

An effort should be made to keep pipetted samples free of gross debris. It is recommended that highly turbid specimens be centrifuged before analysis. Adulteration of the urine sample may cause erroneous results. If adulteration is suspected, obtain another sample and forward both specimens to the laboratory for testing.

**Handle all urine specimens as if they were potentially infectious.**

### Assay Procedure

Analyzers capable of maintaining a constant temperature, pipetting samples, mixing reagents, measuring enzymatic rates at 340 nm and timing the reaction accurately can be used to perform this assay.

Before performing the assay, refer to the analyzer-specific protocol sheet, which contains parameters and/or additional instructions for use.

### Quality Control and Calibration

Good laboratory practice suggests the use of control specimens to ensure proper assay performance. Use controls near the cutoff calibrator to validate the calibration. Control results must fall within established ranges as determined by your laboratory. If results fall outside of established ranges, assay results are invalid.

### Qualitative Analysis

For qualitative analysis of samples, use the 200 ng/mL calibrator as a cutoff level. The DRI® MultiDrug Urine Calibrator 2, which contains 200 ng/mL secobarbital, is used as a cutoff reference for distinguishing "positive" from "negative" samples. In certain applications, 300 ng/mL has been used as a cutoff calibrator.

### Semiquantitative analysis

For semiquantitative analysis, use all calibrators.

### Results and Expected Values

#### Qualitative results

A sample that exhibits a change in absorbance ( $\Delta A$ ) value equal to or greater than the value obtained with the cutoff calibrator is considered positive. A sample that exhibits a change in absorbance ( $\Delta A$ ) value lower than the value obtained with the cutoff calibrator is considered negative.

#### Semiquantitative results

A rough estimate of drug concentration in the samples can be obtained by running a standard curve with all calibrators and quantitating samples off the standard curve.

## Limitations

1. A positive result from this assay indicates only the presence of barbiturates and does not necessarily correlate to the extent of physiological and psychological effects.
2. A positive result by this assay should be confirmed by another nonimmunological method such as GC, TLC or GC/MS.
3. The test is designed for use with human urine only.
4. It is possible that other substances and/or factors (eg, technical or procedural) not listed in the specificity table may interfere with the test and cause false results.

## Specific Performance Characteristics

Typical performance data results obtained on the Hitachi 717 analyzer are shown below.<sup>4</sup>

### Precision

The Negative, 200 ng/mL calibrator, 1000 ng/mL calibrator, Control 1 and Control 2 were assayed, and the following results were obtained:

### Qualitative

Calibrator	Within-run (n=20)		Run-to-run (n=17)	
	Mean ± SD (mA/min)	% CV	Mean ± SD (mA/min)	% CV
0	147 ± 1.4	0.9	147 ± 0.7	0.5
200	239 ± 1.2	0.5	235 ± 0.8	0.3
1000	340 ± 2.1	0.6	332 ± 1.5	0.5

### Semiquantitative

Control	Within-run (n=20)		Run-to-run (n=17)	
	Mean ± SD (ng/mL)	% CV	Mean ± SD (ng/mL)	% CV
Control 1	157 ± 1.4	0.9	161 ± 2.2	1.4
Control 2	264 ± 1.2	0.8	264 ± 3.3	1.3

### Sensitivity

Sensitivity, defined as the lowest concentration that can be differentiated from the negative urine calibrator with 95% confidence, is 25 ng/mL.

### Accuracy

One hundred and four clinical urine specimens were tested with a commercially available EIA assay and DRI Barbiturate Assay. There was 100% agreement between the two methods. Seventy-eight samples were positive and twenty-two were negative by both assays. In addition, all seventy-eight positive samples were confirmed positive by the GC/MS method.

## Specificity

Various potentially interfering substances were tested for cross-reactivity with the assay. The compounds listed in the table below produced a result approximately equivalent to the cutoff calibrator.

Compound	Concentration Tested (ng/mL)
Alphenal	250
Amobarbital	200
Aprobarbital	200
Barbital	1500
Butabarbital	250
Butalbital	300
Butethal	300
Diallybarbital	600
Pentobarbital	500
Phenobarbital	600
Secobarbital	200
Talbutal	60
Thiopental	600

The compounds listed in the table below produced a negative result relative to the cutoff calibrator.

Compound	Concentration Tested (µg/mL)
Acetaminophen	1000
Acetylsalicylic acid	1000
d-Amphetamine	1000
Benzoyllecgonine	1000
Caffeine	100
Codeine	1000
Hydroxyphenytoin (HPPH)	500
Meperidine	1000
Methadone	1000
Methaqualone	1000
Morphine	1000
Oxazepam	500
Phencyclidine	1000
Phenytoin (DPH)	500
Propoxyphene	1000

## References

1. *Urine Testing for Drug of Abuse*. National Institute on Drug Abuse (NIDA) Research Monograph 73, 1986.
2. *Mandatory Guidelines for Federal Workplace Drug Testing Program*. National Institute on Drug Abuse. Federal Register Vol. **53**, No 69, pp 11970 (1988).
3. Rubenstein KE, Schneider RS, and EF Ullman: *Homogeneous enzyme immunoassay: a new immunochemical technique*. Biochem Biophys Res Commun 47:846-851 (1972).
4. Data on file at Microgenics Corporation.

### Manufacturer:

Microgenics Corporation  
46360 Fremont Blvd.  
Fremont, CA 94538 USA  
US Toll Free: 1-800-232-3342



### Authorized Representative in E.U.:

Microgenics GmbH  
Spitalhofstrasse 94  
D-94032 Passau Germany  
Tel: +49 (0) 851 886 89 0  
Fax: +49 (0) 851 886 89 10

### Other countries:

Please contact your local Microgenics representative.